

Implant Compatibility Chart

Manufacturer	Implant	Custom Abutment			Ti base	Lab Analog	Lab Screw	Screw Torque
		Titanium	Zirconia	Hybrid (Ti base + cementable zirconia)				
3I	Certain3.4	●		●	●	●	●	20
	Certain4.1	●		●	●	●	●	20
	Certain5.0	●		●	●	●	●	20
	Certain6.0	●		●	●	●	●	20
Zimmer	Tapered Screw-vent 3.5	●		●	●	●	●	30
	Tapered Screw-vent 4.5	●		●	●	●	●	30
	Tapered Screw-vent 5.7	●		●	●	●	●	30
BioHorizons	Internal3.0	●		●	●		●	30
	Internal3.5	●		●	●		●	30
	Internal4.5	●		●	●		●	30
	Internal5.7	●		●	●		●	30
Astra	OsseoSpeed 3.5/4.0	●		●	●	●	●	25
	OsseoSpeed 4.5/5.0	●		●	●	●	●	25
Nobelbiocare Branemark	NP	●	●	●	●	●	●	35
	RP	●	●	●	●	●	●	35
	WP	●	●	●	●	●	●	35
Nobelbiocare Replace	NP	●		●	●	●	●	35
	RP	●		●	●	●	●	35
	WP	●		●	●	●	●	35
Nobelbiocare Active	NP	●		●	●	●	●	35
	RP	●		●	●	●	●	35
Straumann (ITI)	NN	●		●	●	●	●	35
	RN	●		●	●	●	●	35
	WN	●		●	●	●	●	35
Straumann Bone Level	NC (Φ3.3)	●		●	●	●	●	35
	RC (Φ4.1/Φ4.8)	●		●	●	●	●	35
SIC	3.3	●		●	●	●	●	20
	4.2	●		●	●	●	●	20
Friadent	Frialit 3.0	●		●	●	●	●	24
	Frialit 3.4	●		●	●	●	●	24
	Frialit 3.8	●		●	●	●	●	24
Xive	Frialit 4.5	●		●	●	●	●	24
	Frialit 5.5	●		●	●	●	●	24
	Frialit 6.5	●		●	●	●	●	24
Ankylos	Regular C	●		●	●	●	●	15
Camlog	Φ3.3	●		●	●		●	20
	Φ3.8	●		●	●		●	20
	Φ4.3	●		●	●		●	20
	Φ5.0	●		●	●		●	20
Bego	Splic1.75	●		●	●		●	30
	Splic3.25	●		●	●		●	30

	Splie4.1	•		•	•		•	30
	Splie4.5	•		•			•	30
Osstem	GS Mini (Φ3.5)	•		•	•	•	•	30
	GS Standard (Φ4.0/Φ4.5/Φ5.0)	•		•	•	•	•	30
Dentium	DAN38	•		•	•		•	30
Dio(SM Internal)	SM Internal Na 3.8	•		•	•		•	
	SM Internal Re-WI(4.5 5.3)	•		•	•		•	
Anthogyr	Axiom	•		•	•		•	25
MIS	Internal Hex Narrow (Φ3.3)	•		•	•			35
	Internal Hex Standard (Φ3.75, Φ4.20)	•		•	•			35
	Internal Hex Wide (Φ5.0, Φ6.0)	•		•	•			35

Turnaround times

In-Lab Turnaround times <i>(abutment only)</i>		
Titanium		Zirconia
2		2

Note: Turnaround time listed above only applies when the custom abutment is manufactured with another Mabel Dental Lab product. If you only want a custom abutment manufactured, please allow for a total of 8 business days. The time listed begins the first business day after receipt of all case information and material required and does not include shipping, weekends or holidays.

Other services

Service	Additional turnaround days
Crown/Bridge	add 6 days
Removable diagnostic preparation	add 6 days

- We do not provide any implant parts unless otherwise specified. We recommend to all of our Smile Partners to provide us with all of the necessary implant parts in order to save time and money. We can order any implant parts as necessary, however we need the implant dimensions and manufacturer and require up to 3 additional days when ordering parts.